



Certificate of Analysis

ISO 9001

Reference Material

Product name

4-(2-Hydroxyethyl)indolin-2-one

Product code

MM0331.08-0025

CAS number

139122-19-3

Molecular weight

177.20

Molecular formula

C₁₀H₁₁NO₂

Lot number

1149538

Appearance


beige solid

Melting point (DSC)

149 °C

Long-term storage

2 to 8 °C, dark



Assay "as is"
96.7 %

Date of shipment:

08 Nov 2021

Producer confirms that this reference material (RM) meets the specification detailed on this Certificate of Analysis for **two years** from the date of shipment, provided the substance is stored under the recommended conditions unopened in the original container.

Release by:	Date of Release:		Product Release
Dr. Sabine Schröder	Luckenwalde, 07 Oct 2021		



Mikromol™

Product information

For laboratory use only. Not suitable for human or animal consumption.

Before usage of the RM, it should be allowed to warm to room temperature. No drying required, as the certified value is already corrected for the content of water and other volatile materials.

The product quality is controlled by regularly performed quality control tests (retests).

Further content

Identity

Assay


Final result

Revision table

SPECIMEN




Method	Conditions	Result
¹³ C-NMR	100 MHz, DMSO-d ₆	Structure confirmed



SPECIMEN




Method	Conditions	Result
MS	3.5 kV ESI+; capillary temperature: 269 °C Theoretical value: 178.08626	Structure confirmed



SPECIMEN



Method	Conditions	Result
IR	Attenuated Total Reflection Fourier Transform Infrared (ATR-FTIR) Spectroscopy	Structure confirmed



Assay


The assay of the reference material was assessed by following analyses.

Purity by high performance liquid chromatography (HPLC)

HPLC Conditions:	
Column	Cortecs UPLC C18 +; 1.6 μ m, 75 x 2.1 mm
Column temperature	40 °C
Detector	DAD, 200 nm
Injector	Auto 1 μ l; 0.079 mg/ml in Acetonitrile/Water 50/50 (v/v)
Flow rate	0.5 ml/min
Phase A	Water, 0.1 % HCOOH
Phase B	Acetonitrile, 0.1 % HCOOH
Gradient program	0-1 min A/B 98/2 1-4 min A/B to 2/98 4-5 min A/B to 98/2 5-9 min A/B 98/2 (v/v)



HPLC chromatogram and peak table



Area percent report - sorted by signal

Pk #	Retention time	Area	Area %
1	3.389	16.8119	99.60
2	3.849	0.0378	0.22
3	4.535	0.0304	0.18
Totals		16.8801	100.00

The content of the analyte was determined as ratio of the peak area of the analyte and the cumulative areas of the purities, added up to 100 %. System peaks were ignored in calculation.

Result (n = 3)

99.60 %; SD = 0.01 %



Volatile content

Water content	
Method	Karl Fischer titration
Result (n = 3)	2.82 %; SD = 0.22 %

Residual solvents	
Method	¹ H-NMR
Result (n = 1)	Sum: 0.07 % 0.07 % Methanol

Final result

Assay "as is": 96.72 %

The assay "as is" is assessed by 100% method (mass balance) and is equivalent to the assay based on the not anhydrous and not dried substance respectively.

The calculation of the 100% method follows the formula:

$$\text{Assay (\%)} = (100\% - \text{volatile contents (\%)}) * \frac{\text{Purity (\%)}}{100\%}$$

Volatile contents are considered as absolute contributions and purity is considered as relative contribution. Inorganic residues are excluded by additional tests.

Revision table

Revision	Date	Reason for revision
00	07 Oct 2021	Release of the Certificate of Analysis - initial version

Product warranties for the RM are set out in the terms and conditions of purchase.