

Alcan Research and
Development Centre

Centre de recherche et
de développement Arvida

Alcan Non-metallic Standards
Certificate of Analysis

Échantillons-témoins
non métalliques Alcan
Certificat d'analyse

Material Electrolytic Bath

Identification BA-02

	Ratio	Excess AlF ₃ (%)	α Al ₂ O ₃ (%)	Free Alumina (%)	CaF ₂ (%)
Assigned value *	1.12	10.8	0.3	2.7	5.7
Standard deviation	0.007	0.3	0.09	0.1	0.03
Number of values	50	50	40	10	50

* Values are not corrected for α Al₂O₃ content.

Certified by D. Bussières


Luc Duchesneau
Responsible for certification

18 November 1987