

Gravimetric Certificate



ISO Guide 34 Certified Reference Material

Product Identification

Article Code: DRE-L11322000AL
Article Name: Chloridazon-desphenyl
Formula: C₄H₄CIN₃O
Mol. Weight: 145.55
CAS No.: 6339-19-1

Lot Number: H170885AL
Expiry Date: 06.04.2022
Storage Temperature: 20°C ± 4°C

Storage and handling: The CRM should be stored in the original sealed bottle at the temperature given above. After use the bottle should be tightly closed and protected from moisture and light. The expiry date is valid for original sealed bottles under recommended storage conditions only.

Gravimetric Data

Compound Name: Chloridazon-desphenyl	Lot. No.: G145307	Purity: 99.2 %	Weight: 10.078 mg
Batch Solvent: Acetonitrile	Solvent Lot: 17057042	Batch size: 1000.00 ml	

Concentration: 10.00 mg/l Expanded Uncertainty U: 0.25 mg/l

The uncertainty of this standard is calculated in accordance with the ISO Guide 34 and EURACHEM/CITAC Guide - Quantifying Uncertainty in Analytical Measurement, Second Edition. The expanded uncertainty is $U(\text{exp}) = u(\text{CRM}) \times k$, where k is the coverage factor at the 95% confidence level ($k=2$). Uncertainty $u(\text{CRM})$ is based on the combination of the uncertainties associated with each individual operation involved in the analysis of the product: $u(\text{CRM}) = \sqrt{u(\text{char})^2 + u(\text{bb})^2 + u(\text{lbs})^2 + u(\text{sts})^2}$; $u(\text{char})$ is the uncertainty of characterisation; $u(\text{bb})$ uncertainty of homogeneity test; $u(\text{lbs})$ uncertainty of stability test long-term; $u(\text{sts})$ uncertainty of stability test short-term. $u(\text{lbs})$ and $u(\text{sts})$ are not included in the calculation as the stability statement is based on real evidence opposed to simulation.

Minimum sample: 1 ml is recommended as the minimal sample amount. If less material is used, it is recommended to increase the certified uncertainty by a factor of two for half sample and a factor of four for a quarter of sample.

Intended use: Use this CRM as calibrant for chromatography or any other analytical technique.

Analytical Data

Traceability of chromatography: To the International System of Units (SI).

Instrument: HPLC/DAD
Detection: DAD
Column: Lichrospher 100 Diol 5 µm 250 x 4 mm
Inj.-Vol.: 10 µl
Flow: 1.0 ml/min
Ret.Time: 3.45 min

Method Details
Acetonitrile:Water+0.5% H₃PO₄ 9:1

Comment

Traceability: The balances used are calibrated with weights traceable to the national standards (DKD).
Calibrated class A glassware is used for volumetric measurements.

Certificate Revision 1 - 06.04.2018 - D. Schmid

Certified on: 06.04.2018
Certified by: D. Schmid
RM Release


Schmid D.

The LGC Labor GmbH, accredited by DAkkS as indicated by the accreditation number D-RM-19883-01 & D-PL-19883-01, has shown competence based on ISO Guide 34:2009 with relevant parts of DIN EN ISO/IEC 17025:2005 for production of certified reference materials in form of organic pure substances and in form of single and multi-component solutions of organic pure substances.

LGC Labor GmbH - Bgm.-Schlosser-Straße 6A - 86199 Augsburg - Germany
Phone +49 821 906080 - Fax +49 821 9060888 - augsburg.inquiry@lgcgroup.com
The warranty for this product is limited to the purchasing price of this product.

Data file:	11322000-05.dx	Instrument:	DAD5
Sample name:	H170885AL	Sequence Name:	06032018-1
Inj. volume [µl]:	10.0	Injection date:	3/6/2018 3:57:37 PM
Acq. method:	S5_91PK.amx	Location:	P1-C1

Sample Description Chloridazon-desphenyl
 Column: Lichrospher 100 Diol 5µm 250 x 4 mm



Signal: * DAD1A,Sig=205,2 Ref=off

Nr.	RT [min]	Area	Height	Area%
1	3.45	586.24892	126.32	100.00
	Sum	586.25		

Schmid D.