



CERTIFIED REFERENCE MATERIAL BCR[®] – 339

CERTIFICATE OF ANALYSIS

6H-BENZO[c,d]PYREN-6-ONE ¹⁾		
	Mass fraction	
	Certified value ²⁾ [g/g]	Uncertainty ³⁾ [g/g]
6H-Benzo[c,d]pyren-6-one	0.988	0.009
1) CAS # 3074-00-8. 2) This value is the unweighted mean of 24 sets of results, independently obtained from 8 laboratories. The value is traceable to the SI. 3) The certified uncertainty is the expanded uncertainty estimated in accordance with the Guide to the Expression of Uncertainty in Measurement (GUM) with a coverage factor $k = 2$, corresponding to a level of confidence of about 95 %.		

This certificate is valid for one year after purchase.

Sales date:

The minimum amount of sample to be used is 0.1 mg.

NOTE

This material has been certified by BCR (Community Bureau of Reference, the former reference materials programme of the European Commission). The certificate has been revised under the responsibility of IRMM.

Brussels, March 1990
Revised: December 2007

Signed: _____

Prof. Dr. Hendrik Emons
Unit for Reference Materials
EC-JRC-IRMM
Retieseweg 111
2440 Geel, Belgium

DESCRIPTION OF THE SAMPLE

The material is available in amber glass vials containing 10 mg.

ANALYTICAL METHOD USED FOR CERTIFICATION

- Gas chromatography (GC)
- High performance liquid chromatography (HPLC)
- Gas chromatography/Mass spectrometry (GC/MS)
- Mass spectrometry

PARTICIPANTS

- Biochemisches Institut für Umweltcarcinogene, Ahrensburg (DE)
- Bundesanstalt für Materialprüfung, Berlin (DE)
- Centre d'Etudes et de Recherches des Charbonnages de France, Verneuil (FR)
- CNR Istituto Inquinamento Atmosferico, Roma (IT)
- Energie Centrum Nederland (ECN), Petten (NL)
- Instituut voor Toegepaste Chemie, TNO, Zeist (NL)
- Istituto Superiore di Sanità, Roma (IT)
- National Physical Laboratory, Teddington (GB)
- Instituut voor Toegepaste Chemie, TNO, Delft (NL)
- Risø National Laboratory, Roskilde (DK)
- Studiecentrum voor Kernenergie/Centre d'Etude de l'Energie Nucléaire, Mol (BE)
- Union Technique de l'Automobile, du Motorcycle et du Cycle (UTAC), Montlhéry (FR)

SAFETY INFORMATION

Cancer suspect agent. The material must be handled with great care, especially avoiding skin contamination, ingestion or inhalation. Discard solutions after use in accordance with appropriate safety regulations for carcinogenic or cancer suspect agents.

INSTRUCTIONS FOR USE

This material is intended mainly for the qualitative and quantitative calibration of analytical apparatus and methods and for the study of biological activity. Solutions for calibration purposes should be freshly prepared and be protected from extended exposure to light. Discard solutions after use.

STORAGE

The material should be stored in the dark at 4 °C to prevent photo-oxidation reactions. However, the European Commission cannot be held responsible for changes that happen during storage of the material at the customer's premises, especially of opened samples.

LEGAL NOTICE

Neither IRMM, its subsidiaries, its contractors nor any person acting on their behalf,

(a) make any warranty or representation, express or implied that the use of any information, material, apparatus, method or process disclosed in this document does not infringe any privately owned intellectual property rights;

or

(b) assume any liability with respect to, or for damages resulting from, the use of any information, material, apparatus, method or process disclosed in this document save for loss or damage arising solely and directly from the negligence of IRMM or any of its subsidiaries.

NOTE

A technical report on the production of BCR-339 is available on the internet (<http://www.irmm.jrc.be>). A paper copy can be obtained from IRMM on request.