

**Certified Reference Material**

This certificate is designed in accordance with ISO 17034 and ISO Guide 31. This certified reference material (CRM) was designed, produced and verified in accordance with ISO/IEC 17025, ISO 17034 and a registered quality management system ISO 9001.

Product Name

Dimethachlor-oxalamic acid (OA) 100 µg/mL in Acetonitrile

Product Code

DRE-A12670400AL-100

Lot Number

2-H412188AL

CAS No.

1086384-49-7

Format

Solution

Mol. Weight

251.28

Expiry


13 Dec 2021

Mol. Formula

C₁₃H₁₇NO₄

Storage

≤ -10 °C



CERTIFIED
Concentration
99.96 µg/m


CERTIFIED
Expanded Uncertainty (U)
5.5 µg/mL

Uncertainty

The certified value(s) and uncertainty(ies) are determined in accordance with ISO 17034 with an 95% confidence level (k=2). Uncertainty is based on the Total Combined Uncertainty, including uncertainties of preparation, purity of neat materials, homogeneity, long-term stability testing, and transportation stability.

The producer certifies that this reference material meets the specification stated in this certificate until the expiry date, provided it is stored unopened at the recommended temperature herein. Product warranties for this reference material are set out in the terms and conditions of purchase.

CERTIFIED BY	CERTIFIED ON		RM Release
HuiChen Stavros, Ph.D.	22 Jun 2020		



Instrument
LC/HRMS - Positive Mode

Column/Flow
Vanquish C18+ 100mm x 2.1mm
ID 1.5um Particle / 0.2 mL/min

Method Details
Mobile Phase A: Water w/0.1%
Formic Acid
Mobile Phase B: Acetonitrile w/
0.1% Formic Acid

Time	%A	%B
0.0	95	5
1.0	95	5
22	5	95
22.25	5	95
28.5	95	5

Method of Preparation

The certified value is based on gravimetric and volumetric preparation of this CRM. This CRM has been confirmed by the appropriate analytical techniques.

Batch Information

Solvent: Acetonitrile, Lot no. 198519, 100 mL

Gravimetric Data

Compound Name	Lot No.	Weight (mg)	Chemical Purity
Dimethachlor Oa	6748.551.1P	10.20	98

Intended Use

This CRM is intended for use in a laboratory as a calibration and quality control standard or in method development for analytical techniques.

Safety

Proper precautions should be observed while handling. See Safety Data Sheet.

Traceability

The balances used for gravimetric measurements are calibrated with weights traceable to the national standards (NIST). The calibration of

the balances is verified daily internally and annually by an external accredited calibration service. Chromatographic methods are traceable to the International System of Units (SI).

Homogeneity

Random replicate samples of the final packaged CRM have been analysed to prove homogeneity compliant with ISO 17034.

Storage

The CRM should be stored in the original sealed container at the indicated temperature.

Instructions for use

The CRM should be used shortly after opening to avoid concentration changes due to evaporation. It is recommended to use 1 µL as the minimum sample size. If storage after opening is necessary, it should be transferred to an amber vial with minimum head space and a Teflon-lined silicon septum. If handled as recommended, use period after opening is a maximum of 550 days for an estimated 5% drift in concentration as a result of analyte and/or solvent transpiration. Visit the support section of our website lgcstandards.com for a series of Dr. Ehrenstorfer Tech Tip videos and frequently asked questions.

LGC Group
7290-B Investment Drive
North Charleston, SC 29418
United States
T | +1 843 763 4884
F | +1 866 509 5146
E | dr.ehrenstorfer@lgcgroup.com

The producer of this reference material is registered to ISO 9001:2015 under IZ391-IS4 by NSF-ISR and accredited to ISO 17025:2005 and ISO 17034:2016 by A2LA with the accreditation numbers 3031.01 and 3031.02.



ISO 17034 Accredited
Reference Material Producer
Cert. No. 3031.02