

CERTIFICATE OF ANALYSIS

511X G6023 (batch A)

Certified Reference Material Information

Type: ALUMINIUM / MAGNESIUM ALLOY (CAST)

Form and Size: Disc 50mm diameter x 17mm thick

Produced by: Coleshill Laboratories Ltd

Certified and Supplied by: MBH Analytical Ltd

Assigned Values

Percentage element by weight

Element	Cu	Mg	Si	Fe	Mn	Ni	Zn
Value ¹	0.451	0.512	1.21	0.51	0.556	0.036	0.0260
Uncertainty ²	0.012	0.006	0.03	0.02	0.010	0.002	0.0009

Element	Pb	Sn	Ti	Cr	Bi	Zr	Be
Value ¹	0.0110	1.037	0.0117	0.0054	0.353	0.0110	0.0020
Uncertainty ²	0.0009	0.010	0.0010	0.0003	0.007	0.0013	0.0002

Definitions

- ¹ The certified values are the present best estimates of the true content for each element. Each value is a panel consensus, based on the averaged results of an interlaboratory testing programme, detailed on page 3.
- ² The uncertainty values are generated from the 95% confidence interval derived from the wet analysis results, in combination with a statistical assessment of the homogeneity data, as described on page 2.

Certified by:

MBH ANALYTICAL LIMITED


C Eveleighon 7th April 2010

Method of Preparation

This reference material was produced from commercial-purity aluminium. The main alloys and trace elements were added as pure elements or master alloys. The melt was cast by sequential transfer of aliquots into individual iron moulds. Approximately 2mm has been removed from the cast face of each disc, to minimise any surface effects.

Sampling

Samples for chemical analysis were taken from various positions throughout the batch. Approximately 10% of all discs were selected for non-destructive homogeneity testing.

Homogeneity

The discs were checked for sample and batch uniformity using an optical emission spectrometer.

Using the combined data for each surface, standard deviation values were derived for each element as an indicator of any non-homogeneity (as determined for the specific sample size taken by the spectrometer).

Chemical Analysis

Analysis was carried out on millings taken from samples representative of the product. It was performed by a panel of laboratories mostly operating within the terms of EN ISO/IEC 17025 - 2005, using documented standard reference methods and validated by appropriate reference materials. The individual values listed overpage are the average of each analyst's results.

Estimation of Uncertainties

Each element certified has been analysed by several laboratories, and 95% half-width confidence intervals ($C_{(95\%)}$) for the resultant mean values have been derived by the method shown on page 3.

As a separate exercise, the degree of non-homogeneity of the batch for each element has been quantified by a programme of non-destructive application testing, discussed above.

The final certified uncertainty for each element has been derived by combining these two factors, using the square-root of the summed squares.

Traceability

Much of the analytical work performed to assess this material has been carried out by laboratories with proven competence, as indicated by their accreditation to ISO 17025. It is an implicit requirement for this accreditation that analytical work should be performed with due traceability, via an unbroken chain of comparisons, each with stated uncertainty, to primary standards such as the mole, or to nationally- or internationally-recognised reference materials. In addition, some of the results derived as part of this testing programme have traceability to NIST standards, as part of the analytical calibration or process control.

Usage

Intended use: With optical emission and X-ray fluorescence spectrometers.

Recommended method of use: Aluminium alloys are generally prepared by machining on a mill or a lathe. However, users are recommended to follow the calibration and sample preparation procedures specified by the relevant instrument manufacturer.

Preparation should be the same for reference materials and the samples for test.

A minimum of five consistent replicate analyses is recommended to provide the necessary sample size. Users are advised to check against possible bias between reference materials and production samples due to differences in metallurgical history, and be aware of possible inter-element effects.

Analytical Data

Percentage element by weight

Sample	Cu	Mg	Si	Fe	Mn	Ni	Zn
1	0.436	0.505	1.163	0.493	0.542	0.0343	0.0245
2	0.438	0.508	1.175	0.498	0.544	0.0348	0.0246
3	0.439	0.508	1.192	0.505	0.548	0.0349	0.0252
4	0.442	0.510	1.198	0.505	0.552	0.0354	0.0255
5	0.449	0.510	1.210	0.507	0.554	0.0368	0.0256
6	0.450	0.511	1.223	0.510	0.557	0.0368	0.0256
7	0.451	0.512	1.225	0.513	0.557	0.0369	0.0257
8	0.452	0.517	1.230	0.520	0.560	0.0372	0.0262
9	0.455	0.520	1.255	0.522	0.561	0.0372	0.0268
10	0.470	0.521		0.524	0.564	0.0382	0.0281
11	0.470				0.564		0.0286
12					0.565		
Mean	0.451	0.512	1.21	0.510	0.556	0.0363	0.0260
Std Dev	0.011	0.005	0.03	0.010	0.008	0.0013	0.0013
C_(95%)	0.008	0.004	0.02	0.007	0.005	0.0009	0.0009

Sample	Pb	Sn	Ti	Cr	Bi	Zr	Be
1	0.0085	1.029	0.0105	0.0049	0.338	0.0083	0.0017
2	0.0089	1.032	0.0109	0.0050	0.341	0.0095	0.0018
3	0.0096	1.033	0.0110	0.0050	0.343	0.0096	0.0019
4	0.0101	1.033	0.0115	0.0052	0.345	0.0096	0.0019
5	0.0103	1.033	0.0116	0.0052	0.351	0.0105	0.0020
6	0.0114	1.037	0.0117	0.0053	0.354	0.0108	0.0020
7	0.0115	1.039	0.0117	0.0055	0.361	0.0116	0.0020
8	0.0118	1.041	0.0121	0.0057	0.361	0.0125	0.0021
9	0.0120	1.042	0.0122	0.0059	0.363	0.0136	0.0022
10	0.0121	1.045	0.0122	0.0061	0.365	0.0137	0.0022
11	0.0122	1.048	0.0127		0.366		
12	0.0130		0.0128				
Mean	0.0110	1.037	0.0117	0.0054	0.353	0.0110	0.0020
Std Dev	0.0014	0.006	0.0007	0.0004	0.010	0.0018	0.0002
C_(95%)	0.0009	0.004	0.0004	0.0003	0.007	0.0013	0.0001

Note: $C_{(95\%)}$ is the 95% half-width confidence interval derived from the equation:

$$C_{(95\%)} = (t \times SD) / \sqrt{n}$$

where n is the number of available values, t is the Student's t value for n-1 degrees of freedom, and SD is the standard deviation of the test results.

Participating Laboratories

Exova Materials Testing Ltd
Universal Scientific Laboratory Pty
Laboratory Testing, Inc
Genitest Inc
Institute of Iron & Steel Technology
Luo Yang Copper Co
South-West Aluminium Group
Sargam Metals Pvt Ltd
TCR Engineering Servs Pvt Ltd
Raghavendra SpectroMet Laboratory
De Bruyn Spectroscopic Solutions
London & Scandinavian Met. Co
Coleshill Laboratories Ltd

Middlesbrough, England
Milperra, NSW, Australia
Hatfield, PA, USA
Montreal, Canada
Shanghai, China
Luo Yang, He Nan, China
Jiulong Puo, Sichuan, China
Chennai, India
Mumbai, India
Bangalore, India
Johannesburg, South Africa
Rotherham, England
Birmingham, England

UKAS accreditation 0239
NATA accreditation 0492
A2LA accreditation 0117
PRI accreditation 123077
CNAL accreditation 0783
CNAL accreditation 0173
CNAL accreditation T007
NABL accreditation T025
NABL accreditation T367
NABL accreditation T371

Note: to achieve the above accreditation (eg UKAS, NATA, etc), test houses are required to demonstrate conformity to the general requirements of EN ISO/IEC 17025.

Analytical Methods Used

ELEMENT	RESULT No. & METHOD		
	ICP-AES	FAAS	OTHER
Copper	1, 3-5, 9, 11	2, 6-8, 10	
Magnesium	5-8, 10	1-4, 9	
Silicon	8	-	1-5, 7 gravimetric (perchloric acid) 6, 9 photometric (molybdenum blue)
Iron	3, 4, 7, 8, 10	1, 5, 9	2, 6 photometric (orthophenanthroline)
Manganese	1-6, 8, 12	7, 10	9, 11 photometric (periodate)
Nickel	2-5, 8-10	1, 6, 7	
Zinc	1, 6-11	2-5	
Lead	1, 3, 4, 6, 9-11	2, 5, 7, 8, 12	
Tin	1-4, 6, 9, 11	5, 7	8, 10 photometric (phenylfluorone)
Titanium	1, 4, 5, 7-9, 11	2, 10	3, 6, 12 photometric (di-antipyryl methane)
Chromium	1, 2, 7-10	3-6	
Bismuth	1, 4-6, 10, 11	2, 3, 8, 9	7 photometric (brucine)
Zirconium	2, 3, 5-10		1 ICP-MS 4 photometric (xylenol orange)
Beryllium	1-3, 6, 7, 9, 10	5, 8	4 photometric (eriochrome cyanine R)

Notes

This Certified Reference Material has been produced and certified in accordance with the requirements of ISO Guide 34-2000, ISO Guide 31-2000 and ISO Guide 35-2006, taking into account the requirements of the ISO Guide to the Expression of Uncertainty in Measurement (GUM).

The unidirectional solidification effects associated with semi-chill casting of this alloy, have led to the formation of inhomogeneous segregates in the rear portion of the disc. The above certification is therefore only applicable from the front face of the disc for a depth of 10mm. Material to the rear of the disc, to a depth of ~7mm, is not certified.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture. All production records will be retained for a period of 20 years from the date of original analysis. This certification will therefore expire in April 2030, although we reserve the right to make changes as issue revisions, in the intervening period.

This sample is also available in the form of chippings.

The specification, preparation, analysis and certification of this product were supervised by C Eveleigh, PhD, Technical Director, MBH Analytical Ltd.

The material to which this certificate of analysis refers is supplied subject to our general conditions of sale.