

Certificate of Analysis

Characterisation methods are accredited according to

ISO 17025

Reference Material

Product name

(6S)-4,5,6,7-Tetrahydro-1,3-benzothiazole-2,6-diamine

Product code Lot number MM0332.02-0025 W1125141 CAS number Appearance 106092-09-5 white solid

Molecular weight Melting point (DSC)

169.25 230 °C

v

Assay¹ "as is" **99.95** % Uncertainty² U **0.2** %

Intended Use: Use for identification and quantification. The assay is verified by a second testing method.

Date of shipment: 20 Jul 2021

Producer confirms that this reference material (RM) meets the specification detailed on this Certificate of Analysis for **two years** from the date of shipment, provided the substance is stored under the recommended conditions unopened in the original container.

 H_2N_A

Release by: Date of Release:	0	
Dr. Sabine Schröder Luckenwalde, 04 May 2021	Toia	Product Release

¹ Calibration and verification were carried out using standards traceable to SI-units. The value is expressed on an "as is" basis.

² The uncertainty "U" is the expanded uncertainty of the testing method for the assigned value estimated in accordance with the Guide to the Expression of Uncertainty in Measurement (GUM). It corresponds to a level of confidence of about 95%. Coverage factor k = 2.



Product information

This RM is intended for laboratory use only and is not suitable for human or animal consumption.

This RM conforms to the characteristics of a primary standard as described in the ICH/Guidelines. The values quoted in this Certificate of Analysis are the producer's best estimate of the true values within the stated

uncertainties and based on the techniques described in this Certificate of Analysis. The characterisation of this material was undertaken in accordance with the requirements of ISO/IEC 17025. The identity is verified by data

from international scientific literature.

Storage and handling

Before usage of the RM, it should be allowed to warm to room temperature. No drying is required, as assigned values are already corrected for the content of water and other volatile materials.

Reference Material quality is controlled by regularly performed quality control tests (retests).





Assigned value

Assay "as is":

99.95 %; U = 0.20 %

The assay "as is" is assessed by 100% method (mass balance) and is equivalent to the assay based on the not-anhydrous and not-dried substance. The assay is verified by quantitative NMR spectroscopy.

The verified result lies inside our acceptance criteria, i.e. less than 1.0 % difference to assay assigning technique.

For quantitative applications, use the assay as a calculation value on the "as is basis". The uncertainty of the assay can be used for estimation/calculation of measurement uncertainty.

Method 1: Value assigning technique - 100% method

100% method (mass balance) with chromatographic purity by HPLC

Result

99.95 %; U = Q.20 %

The calculation of the 100% method follows the formula:

Assay (%) = (100% - volatile contents (%))

Purity (%)

Volatile contents are considered as absolute contributions and purity is considered as relative contribution. Inorganic residues are excluded by additional tests.

LGC GmbH, Louis-Pasteur-Str. 30, D-14943 Luckenwalde, Germany

MM0332.02-0025

Page 3/10



Method 2: Value verifying technique - quantitative NMR spectroscopy			
Conditions	400 MHz, CD₃COOD		
Internal Standard	1,2,4,5-Tetramethylbenzene (certified reference material), signal 6.4 - 7,2 ppm, 2 H		
Result (mass fraction, n = 6)	99.21 %		






Purity

Purity by high performance liquid chromatography (HPLC)

HPLC Conditions:	
Column	Hypersil Gold C18; 5 µm, 150 x 4.6 mm
Column temperature	40 °C
Detector	DAD, 264 nm
Injector	Auto 3/µl/ 0.1082/ mg/ml in 4.6g/l C ₈ H ₁₇ NaO ₃ S + 9.1g/l KH ₂ PQ ₄ , pH/3.0
Flow rate	1.0 ml/min
Phase A	4.6gXl C ₈ H ₁ NaO ₃ S + 9.1g/l KH ₂ PO ₄ , pH 3.0
Phase B	(4.6g/l C ₈ H ₁ √NaO ₃ S)+9.1g/l KH ₂ PO ₄ ,pH 3-0)/Acetonitrile 50/50 (v/v)
Gradient program	0-7 min A/B 60/40 7-10 min A/B to 20/80 10-13 min A/B to 60/40 13-20 min A/B 60/40 (v/v)

HPLC chromatogram and peak table




LGC GmbH, Louis-Pasteur-Str. 30, D-14943 Luckenwalde, Germany

Page 5/10 MM0332.02-0025 Lot number W1125141



Area percent report - sorted by signal			
Pk #	Retention time	Area	Area %
1	2.472	0.0117	0.01
2	3.385	0.0065	0.01
3	3.895	0.0273	0.03
4	7.600	81.0848	99.94
Totals		81.1303	100.00

The content of the analyte was determined as ratio of the peak area of the analyte and the cumulative areas of the purities, added up to 100 %. System peaks were ignored in calculation.



Method: Elementary analysis

Inorganic residues can be excluded by elementary analysis (CHN).

Page 6/10 MM0332.02-0025 Lot number W1125141




Identity

The identity is assessed by ISO/IEC 17025 accredited testing methods.







Method	Conditions	Result
¹³ C-NMR	100 MHz, DMSO-d ₆	Structure confirmed











LGC GmbH, Louis-Pasteur-Str. 30, D-14943 Luckenwalde, Germany

Page 10/10 Lot number W1125141