

Certificate of Analysis

Reference Substance

Chlorocresol (4-Chloro-3-methylphenol)


Catalogue Number: LGCFOR0459.00

Lot Number: 55035Molecular Formula: C_7H_7CIO Molecular Weight: 142.58CAS Number: [59-50-7]

Long-term Storage: 2 to 8 °C, dark

Appearance: white solid

Melting Point: 65 °C
Assay 'as is': 99.95 %



Date of shipment: 2016-May-20

This certificate is valid for two years from the date of shipment provided the substance is stored under the recommended conditions.

Release Date: 2014-08-08

Dr. Sabine Schröder Product Release

LGC GmbH








I. Identity

The identity of the reference substance was established by following analyses.

Ia. ¹H-NMR Spectrum

Conditions: 400 MHz, DMSO-d₆


The structure is confirmed with the signals of the spectrum and their interpretation.








Ib. Mass Spectrum



Abundance



m/z	fragments
142	[M]
107	[M – Cl]


The signals of the mass spectrum and their interpretation are consistent with the structural formula.





Ic. IR Spectrum

Method: Attenuated Total Reflection Fourier Transform Infrared (ATR-FTIR) Spectroscopy



The signals of the IR spectrum and their interpretation are consistent with the structural formula.

II. Purity


The purity of the reference substance was analysed by high performance liquid chromatography (HPLC).

HPLC Conditions:

Conditions: Column: **Detector:** Injector: RP 60 Select B 1.0 ml/min, 40 °C DAD Auto - 10 min Water/Acetonitrile 65/35 5 µm, 125 x 4 mm 225 nm 5 µl; 0.12 mg/ml in 10 - 15 min Water/Acetonitrile to 50/50 Water/Acetonitrile 50/50 (v/v) 15 - 20 min Water/Acetonitrile to 65/35 20 - 35 min Water/Acetonitrile 65/35 (v/v); 0.1 % H₃PO₄







Area Percent Report - Sorted by Signal

Pk#	Retention Time	Area	Area %
1	4.81	9571614	100.00
Totals		9571614	100.00

For the calculation the system peaks were ignored. The content of the analyte was determined as ratio of the peak area of the analyte and the cumulative areas of the purities, added up to 100 %.

Results:

Average 100 %

Number of results n=3

Standard deviation < 0.01 %



LGCFOR0459.00 Lot Number 55035



III. Water Content

Method: Karl Fischer titration

Results:

IV. Residual Solvents

Method: 1H-NMR

No significant amounts of residual solvents were detected (< 0.05 %)

V. Final Result

Total impurities (HPLC) 0.00% Water content 0.05%

Residual solvents No significant amounts of residual solvents were detected (< 0.05 %).

Assay (100 % method)¹ 99.95 %

The assay is assessed to be 99.95 % 'as is'

The assay 'as is' is equivalent to the assay based on the not anhydrous and not dried substance respectively.

¹ The calculation of the 100 % method follows the formula:

Assay (%) = (100 % - KF - RES) * $\frac{Purity HPLC (%)}{100 \%}$

Water (KF) and Residual solvents (RES) are considered as absolute contributions, HPLC purity is considered as relative contribution.

Standards

LGCFOR0459.00 Lot Number 55035

Excellence through measurement