

## Certificate of Analysis

European Pharmacopoeia

0.5M Potassium Hydroxide, Alcoholic

*Product No:* 3005000  
*Lot No:* 350013L1  
*Expiry date:* 28/11/2014

*Mean Molarity:* 0.5007 M @20°C  
*Date of measurement:* 04/11/2013

**Specification:**  
0.4975 - 0.5025M

**Method:**  
Tested by potentiometric titration, to in-house method TPATB

**Traceability:**  
This volumetric solution was checked by means of Hydrochloric Acid Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 723d, Tris(hydroxymethyl)aminomethane. All raw materials used to prepare this product are of high purity.

Products are manufactured under an NSAI registered I.S EN ISO9001:2008 Quality System, registration no: 19.2769

*Date of issue of the certificate:* 04/11/2013



**Q.C. Supervisor Margaret Dornan**

Complementary information relative to this product is available at [www.reagecon.com](http://www.reagecon.com)  
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-13K16