

Arvida Research and  
Development Centre

Centre de recherche et  
de développement Arvida

Alcan Non-metallic Standards  
Certificate of Analysis

Échantillons-témoins  
non métalliques Alcan  
Certificat d'analyse

**ELECTROLYTIC BATH**

**BA-06**

	Ratio	Excess AlF <sub>3</sub> (%)	$\alpha$ -Al <sub>2</sub> O <sub>3</sub> (%)	Free Alumina (%)	CaF <sub>2</sub> (%)
Assigned Values*	1.00	12.2	16.7	23.3	4.5
Standard deviation	0.01	0.3	0.2	0.4	0.04
Number of values	50	50	40	10	50

\* Values are not corrected for  $\alpha$ -Al<sub>2</sub>O<sub>3</sub> content.

Certified by D. Bussières  
Revised by G.F. Giasson (July 88)



Luc Duchesneau  
Responsible for Certification  
November 1987