



Dr. Ehrenstorfer
276 Abby Road | Manchester, NH 03103 | USA
Tel: +1-603-206-0759 | Fax: 1-603-622-5180
www.lgcstandards.com

Date Received: _____

Certificate of Analysis

Rev

Catalog No.: DRE- YS09010009ME	Lot No.: 264028	Storage: ≤ -10 °C	Solvent: Methanol	Exp. Date: 3-Nov-2017	Description: 2,6-Dichlorophenol Solution, 1,000 mg/L, 5 x 1 mL (RM, ISO GUIDE 34)	Page 1
---	---------------------------	-----------------------------	-----------------------------	---------------------------------	--	---------------

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
2,6-dichlorophenol	87-65-0	99.9	117.9.1.2P	1011 +/- 10.31 mg/L



Katelynn Hardee

Katelynn Hardee

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence)
values listed are determined gravimetrically.

Manufacturer is ISO/IEC 17025:2005 Accredited by A2LA - Certificate Number 3031.01 for Chemical Testing
Manufacturer is ISO/IEC Guide 34:2009 Accredited by A2LA - Certificate Number 3031.02 for Reference Material Manufacturer
Manufacturer's Quality System Audited & Registered by NSF-ISR to ISO 9001:2008