Certificate of Analysis

ISO Guide 34 Reference Material

Product Identification

Article Code: DRE-C20550000
Article Name: Benzo[b]chrysene
Formula: C22H14
Mol. Weight: 278.35
CAS No.: 214-17-5

Lot Number: G145306
Expiry Date: 14.02.2024
Storage Temperature: 20°C ± 4°C

Storage and handling: The RM should be stored in the original sealed bottle at the temperatur given above. After use the bottle should be tightly closed and protected from moisture and light.
The expiry date is valid for original sealed bottles under recommended storage conditions only.

Purity: 99.78% (g/g)

Expanded Uncertainty U= 0.30% (g/g)

The uncertainty of this standard is calculated in accordance with the ISO Guide 34 and EURACHEM/CITAC Guide - Quantifying Uncertainty in Analytical Measurement, Second Edition. The expanded uncertainty is U(exp) = u(RM) x k, where k is the coverage factor at the 95% confidence level (k=2). Uncertainty u(RM) is based on the combination of the uncertainties associated with each individual operation involved in the analysis of the product: $u(RM) = \sqrt{u(char)^2 + u(bb)^2 + u(lts)^2 + u(sts)^2}$; u(char) is the uncertainty of purity determination; u(bb) uncertainty of homogeneity test; u(lts) uncertainty of stability test long-term; u(sts) uncertainty of stability test short-term. u(lts) and u(sts) are not included in the calculation as the stability statement is based on real evidence opposed to simulation.
Minimum sample: 1 mg is recommended as the minimal sample amount. If less material is used, it is recommended to increase the certified uncertainty by a factor of two for half sample and a factor of four for a quarter of sample.
Intended use: Use this RM as calibrant for chromatography or any other analytical technique.

Analytical Data

Traceability of chromatography: To the International System of Units (SI).

Instrument:	GC/FID	Injector:	320°C
Detection:	FID	Initial Temp:	120°C for 4 min
Column:	Optima-5MS, 0.25 µm, 0.25 mm	End Temp:	320°C for 3 min
Inj.-Vol.:	1 µl	Gradient:	15°C/min
Flow:	1.0 ml/min		
Ret.Time:	21.01 min		

Comment
Traceability: The balances used are calibrated with weights traceable to the national standards (DKD).
Calibrated class A glassware is used for volumetric measurements.

Identity: EA, NMR, RT, IR, UV, MS

Certificate Revision 1 - 14.02.2018 - M. Beck

Certified on: 14.02.2018
Certified by: M. Beck

The LGC Labor GmbH, accreditated by DAkkS as indicated by the accreditation number D-RM-19883-01 & D-PL-19883-01, has shown competence based on ISO Guide 34:2009 with relevant parts of DIN EN ISO/IEC 17025:2005 for production of certified reference materials in form of organic pure substances and in form of single and multi-component solutions of organic pure substances.

LGC Labor GmbH - Bgm.-Schlosser-Straße 6A - 86199 Augsburg - Germany
Phone +49 821 906080 - Fax +49 821 9060888 - augsburg.inquiry@lgcgroup.com
The warranty for this product is limited to the purchasing price of this product.

15.09.2017 MP

Data file:	20550000-31-r001.dx	**Instrument:**	FID 2
Sample name:	70911TO G145306	**Sequence Name:**	2017KW37-2b
Inj. volume [µl]:	1.0	**Injection date:**	9/14/2017 2:31:17 AM
Acq. method:	pahk.amx	**Location:**	74
Sample Description	Benzo[b]chrysene		



Signal: FID1A

Nr.	RT [min]	Area [pA*s]	Height [pA]	Area%	Width [min]
1	19.831	0.53588	0.22	0.22	0.111
2	21.017	248.16684	83.45	99.78	0.265
	Sum	248.70			