

Certificate of Analysis

ISO Guide 34 Reference Material

Product Identification

Article Code: DRE-C10648000

Article Name: Bis-(4-aminophenyl)methane Formula: C13H14N2

Mol. Weight: 198.27 CAS No.: 101-77-9 Lot Number:

G156492

Expiry Date:

06.10.2021

Storage Temperature:

20°C ± 4°C

Storage and handling: The RM should be stored in the original sealed bottle at the temperatur given above. After use the bottle should be tightly closed and protected from moisture and light. The expiry date is valid for original sealed bottles under recommended storage conditions only.

97.52% (g/g)

Expanded Uncertainty U= 0.30% (g/g)

The uncertainty of this standard is calculated in accordance with the ISO Guide 34 and EURACHEM/CITAC Guide - Quantifying Uncertainty in Analytical Measurement, Second Edition. The expanded uncertainty is U(exp) = u(RM) x k, where k is the coverage factor at the 95% confidence level (k=2). Uncertainty u(RM) is based on the combination of the uncertainties associated with each individual operation involved in the analysis of the product: u(RM) = vu(char)² + u(bb)² + u(lts)² + u(sts)²; u(char) is the uncertainty of purity determination; u(bb) uncertainty of homogeneity test; u(lts) uncertainty of stability test long-term; u(sts) uncertainty of stability test short-term. u(lts) and u(sts) are not included in the calculation as the stability statement is based on real evidence opposed to simulation. Minimum sample: 1 mg is recommended as the minimal sample amount. If less material is used, it is recommended to increase the certified uncertainty by a factor of two for half sample and a factor of four for a quarter of samp

Injector:

Initial Temp:

End Temp:

Gradient:

320°C

5°C/min

120°C for 4 min

320°C for 8 min

Intended use: Use this RM as calibrant for chromatography or any other analytical technique.

Analytical Data

Traceability of chrom stography: To the International System of Units (SI).

Instrument: GC/FID Detection:

FID

Optima-5MS, 0.25 µm, 0.25 mm tnj.-Vol.: 1 ul

1.0 ml/min Flow:

Ret.Time: 25.30 min

Traceability: The balances used are calibrated with weights traceable to the national standards (DKD).

Calibrated class A glassware is used for volumetric measurements.

Certificate Revision 1

Water Content: 0.10% (g/g) by Karl-Fischer-Titration (U(exp) = 0.04% (g/g)).

Identity:

EA, NMR, RT, IR, UV; MS

Certified on: Certified by:

06.10.2017 N. Müller

The LGC Labor GmbH, accreditated by DAkkS as indicated by the accreditation number D-RM-19883-01 & D-PL-19883-01, etence based on ISO Guide 34:2009 with relevant parts of DIN EN ISO/IEC 17025:2005 for production of certified reference materials in form of organic pure substances and in form of single and multi-component solutions of organic pure substances. Data file:

10648000-12.dx

Sample name:

71009AL G156492

lnj. volume [µl]:

1.0

Acq. method:

pah.amx

Instrument:

FID 1

Sequence Name:

2017KW40-4b

Injection date:


10/6/2017 1:16:59 AM

Location:

35

Sample Description

Bis-(4-aminophenyl)methane



Signal: Nr.	FID1A	Area [pA*s]	Height [pA]	Area%	Width [min]
	RT [min]				
1	24.099	2.26262	0.76	1.95	0.205
2	25.296	113.38736	36.44	97.62	0.252
3	26.907	0.50345	0.16	0.43	0.044
	Sum	116.15			