

## Certificate of Analysis

### Chemical Oxygen Demand - COD Reagents

### COD Calibration Standard 2000mg/l

<i>Product No:</i>	<b>COD2000</b>
<i>Lot No:</i>	<b>COD2M14M1</b>
<i>Expiry date:</i>	<b>28/12/2016</b>
<i>Mean assay:</i>	<b>1990 mg/l as COD</b>
<i>Date of measurement:</i>	<b>20/01/2015</b>

#### **Specification:**

1990 - 2010mg/l

#### **Method:**

Standardised in accordance with in-house method acidimetric.

#### **Traceability:**

This standard was checked by means of Sodium Hydroxide Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology, NIST SRM 84L, Potassium Hydrogen Phthalate

#### **Reference:**

Vogels Textbook of Quantitive Inorganic Analysis. Fifth Edition.

Products are manufactured under an NSAI registered I.S EN ISO9001:2008 Quality System, registration no: 19.2769

*Date of issue of the certificate:* 20/01/2015

*Derdre Madigan*

**QC Technician *Derdre Madigan***

Complementary information relative to this product is available at [www.reagecon.com](http://www.reagecon.com)  
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-15A2