

CERTIFICATE OF CALIBRATION

ISO 17025 / ISO 17034 CERTIFIED COLOUR REFERENCE STANDARD

Pt-Co/Hazen/APHA (Nominal: 15), Colour

Product Code: PTC0015

Batch Number:

4231002

Certificate No: 4738

Issue Date:

Expiry Date:

Certified Values:

Platinum-Cobalt:

15

± 3

Room temperature: 22.9 °C ± 1 °C

Intended Use:

This product is intended for the calibration and/or verification of equipment used for the measurement of colour of liquid samples.

Certification:

An Agilent Cary 4000 spectrophotometer was used to measure the spectral transmittance of the liquid batch samples from 360 nm to 830 nm at 1 nm intervals. The measurement beam area was a vertical patch, 1 mm wide and 6 mm high centrally positioned through the liquid which was contained in a 50 mm optical glass cell. The CIE chromaticity coordinates were calculated in accordance with ASTM E308 for 2° Observer and Illuminant C. The chromaticity coordinates were converted to a corresponding Yellowness Index in accordance with ASTM E313, and then the equivalent platinum-cobalt value calculated based on ASTM D5386.

Uncertainties of the certified values have been determined in accordance with the "Guide to the Expression of Uncertainty in Measurement" (GUM). All uncertainties are reported as expanded uncertainties derived as the combined standard uncertainty multiplied by a coverage factor of $k=2$, providing a level of confidence of approximately 95 %.

Metrological Traceability:

The accuracy of the instrument is checked periodically using standards calibrated at the National Physical Laboratory. The reference beam was a clear air path. The instrument baseline was carried out with a clear air path in both the reference and sample beam.

Homogeneity & Stability:

This product has been produced according to in-house procedures consistent with ISO 17025 and ISO 17034, and its homogeneity is guaranteed to be fit for purpose when used with a sample size appropriate for the intended measurement method.

Storage & Instructions for handling and use:

The shelf life of this product is guaranteed until the expiry date, provided the bottle is unopened and stored upright between 18 °C to 28 °C. The guarantee is void if the bottle seal is broken. Always keep container sealed when not in use. Follow good hygiene practice. Mix well before use by inversion or gentle shaking. Store in the dark. Use immediately after decanting. Do not return samples to the bottle after use.


Health & Safety:

Please refer to the Safety Data Sheet (SDS) supplied for this product. If you require a copy of the original SDS issued with this product, email paragon.sales@lgcgroup.com quoting product code, batch number and expiry date, which can be found on this certificate. It is the responsibility of the user of this material to establish health & safety practices to determine applicability of regulatory limitations prior to use.



Legal Notice:

Paragon Scientific Limited products are not for any cosmetic, drug, or household application and are to be used only by qualified individuals who are trained in appropriate procedures. No claims against Paragon Scientific Limited of any kind whatsoever, whether based on breach of warranty, alleged negligence, or otherwise, with respect to this product shall be greater than the purchase price. In no event shall Paragon Scientific Limited be liable for any loss of profits or any incidental, special, or consequential damages.



Approved Signatory, Mr. P. Whitehurst, Technical Director

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service (UKAS). It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory (NPL) or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. UKAS is one of the signatories to the Multilateral Agreement of European co-operation for Accreditation (EA) for the mutual recognition of calibration certificates issued by accredited laboratories.

SAMPLE