


20-1500 Methyl Pentadecanoate

Lot no: Q 182:10

Purity: 99+%

GC: Injection : On column
Detector : FID
Column : HP-5 15m x 0.32mm film 0.25um
Carrier : Helium 14psi

Temperature Column:
Initial : 80 C , 3min
Final : 240 C
Rate : 10 C /min
Injector : 240 C
Detector : 270 C



Purity test performed:
Okt 2015
Storage: Dark

Expiry date: Okt 2020