

Certificate of Analysis

Reference Substance

Isosorbide Mononitrate

Catalogue Number:	LGCFOR0502.00	Long-term Storage:	2 to 8 °C, dark
Lot Number:	15759	Appearance:	white solid
Molecular Formula: Molecular Weight:	C ₆ H ₉ NO ₆ 191.14	Melting Point (DSC):	90 °C
CAS Number:	[16051-77-7]	Assay 'as is':	100 %
	O H OH		

Date of shipment: 2016-May-20

0₂N-0

This certificate is valid for two years from the date of shipment provided the substance is stored under the recommended conditions.

 \cap

Н

Release Date: 2012-09-28

LGC GmbH

Dr. Sabine Schröder Product Release





LGC GmbH, Im Biotechnologiepark, TGZ II, D-14943 Luckenwalde, Germany

5 pages




I. Identity

The identity of the reference substance was established by following analyses.

Ia. ¹H-NMR Spectrum

Conditions: 400 MHz, DMSO-d₆

The structure is confirmed with the signals of the spectrum and their interpretation.






LGCFOR0502.00 Lot Number 15759

LGC GmbH, Im Biotechnologiepark, TGZ II, D-14943 Luckenwalde, Germany



Ib. IR Spectrum



The signals of the IR spectrum and their interpretation are consistent with the structural formula.

II. Purity

The purity of the reference substance was analysed by high performance liquid chromatography (HPLC).

HPLC Conditions:


Column:	Conditions:	Detector:	Injector:
Hypersil Gold (C18)	1.0 ml/min, 40 °C	DAD	Auto
5 µm, 150 x 4.6 mm	0 – 10 min Water/Acetonitrile 95/5	210 nm	3 µl; 0.4082 mg/ml in
	10 – 17 min Water/Acetonitrile to 75/25 17 – 25 min Water/Acetonitrile to 95/5 25 – 40 min Water/Acetonitrile 95/5 (v/v);		Water/Acetonitrile 50/50 (v/v)
	0.1 % H ₃ PO ₄		



LGCFOR0502.00 Lot Number 15759

LGC GmbH, Im Biotechnologiepark, TGZ II, D-14943 Luckenwalde, Germany





For the calculation the system peaks were ignored. The content of the analyte was determined as ratio of the peak area of the analyte and the cumulative areas of the purities, added up to 100 %.

Results:Average100 %Number of resultsn=3Standard deviation< 0.01 %</td>



LGCFOR0502.00 Lot Number 15759

LGC GmbH, Im Biotechnologiepark, TGZ II, D-14943 Luckenwalde, Germany



III. Water Content

Method: Karl Fischer titration

No significant amounts of water were detected (< 0.05 %).

IV. Residual Solvents

Method: ¹H-NMR

No significant amounts of residual solvents were detected (< 0.05 %).

V. Final Result

Total impurities (HPLC)	0.00 %
Water content	n. d. (not detected)
Residual solvents	n. d. (not detected)
Assay (100 % method) ¹	100.00 %

The assay is assessed to be 100 % 'as is'

The assay 'as is' is equivalent to the assay based on the not anhydrous and not dried substance respectively.

¹ The calculation of the 100 % method follows the formula:

Assay (%) = (100 % - KF - RES)

Purity HPLC (%) 100 %

Water (KF) and Residual solvents (RES) are considered as absolute contributions, HPLC purity is considered as relative contribution.

LGCFOR0502.00 Lot Number 15759

LGC GmbH, Im Biotechnologiepark, TGZ II, D-14943 Luckenwalde, Germany

Excellence through measurement