

## REFERENCE MATERIAL CERTIFICATE

## Reference Material

This certificate is designed in accordance with ISO Guide 31. This reference material (RM) was designed, produced and verified in accordance with a registered quality management system ISO 9001. All measurements were performed according to ISO/IEC 17025 by a DAkkS accredited laboratory (D-PL-19883-01-00).

## Product Name

Benzyltrimethyloctylammonium chloride

Product Code  
DRE-C10572485

Lot Number  
1050362

CAS No.  
959-55-7


Format  
Neat

Mol. Weight  
283.88

Expiry Date  
16 Dec 2023

Mol. Formula  
 $C_{17}H_{30}N \cdot Cl$

Storage Temp  
 $20^{\circ}C \pm 4^{\circ}C$

**CERTIFIED**

Purity  
98.1% (g/g)

**CERTIFIED**

Expanded Uncertainty (U)  
1.0% (g/g)

## Uncertainty


The certified value(s) and uncertainty(ies) are determined in accordance with EURACHEM/CITAC Guide for "Quantifying Uncertainty in Analytical Measurement, 3rd edition", with an 95% confidence level ( $k=2$ ). Uncertainty is based on the Total Combined Uncertainty, including uncertainties of characterisation and stability testing. Stability values are based on real evidence opposed to simulation.

The producer certifies that this reference material meets the specification stated in this certificate until the expiry date, provided it is stored unopened at the recommended temperature herein. Product warranties for this reference material are set out in the terms and conditions of purchase.

CERTIFIED BY	CERTIFIED ON		RM Release
N. Müller	16 Dec 2019		



## REFERENCE MATERIAL CERTIFICATE



### Instrument

HPLC/DAD

### Column

Nucleodur pi2, 5 µm, 250 x 3 mm

### Method Details

Eluent A: 100% Acetonitrile

Eluent B: Water + H<sub>3</sub>PO<sub>4</sub>

Time [min]	Eluent A [%]	Eluent B [%]
0	10	90
1	10	90
20	100	0
25	100	0

### Flow

1 mL/min

### Method of Characterisation

Purity = 100% – Assay impurities – Water content (KF)

### Method of Identification

EA, NMR, RT, UV, IR, MS

### Batch Information

Water Content: 1.1% (g/g) by Karl-Fischer-Titration (U(exp) = 0.4% (g/g)).

Substance is hygroscopic.

### Intended Use

This RM is intended for use in a laboratory as a calibration and quality control standard or in method development for analytical techniques.

standards (DKD). The calibration of the balances is verified daily internally and annually by an external accredited calibration service. Chromatographic methods are traceable to the International System of Units (SI).

### Instructions for use

It is recommended to use 1 mg as the minimum sample size and if less material is used, to increase the certified uncertainty by a factor of two for half sample and four for a quarter of sample. If storage after opening is necessary, the RM should be tightly closed and kept from light and moisture. If the RM was in a sealed ampoule, it should be transferred to a vial with minimum head space. Visit the support section of our website [lgcstandards.com](http://lgcstandards.com) for a series of Dr. Ehrenstorfer Tech Tip videos and frequently asked questions.

### Safety

Proper precautions should be observed while handling. See Safety Data Sheet.

### Storage

The RM should be stored in the original sealed container at the indicated temperature.

### Traceability

The balances used for gravimetric measurements are calibrated with weights traceable to the national

### LGC Labor GmbH

Bgm.-Schlosser-Straße 6A

86199 Augsburg, Germany

T | +49 821 906080

F | +49 821 9060888

E | [dr.ehrenstorfer@lgcgroup.com](mailto:dr.ehrenstorfer@lgcgroup.com)