

Gravimetric Certificate

**EHRENSTORFER™**

Product Identification

17445500 Thenylchlor
Formula C₁₆H₁₈CINO₂S
Mol.Weight 323.8
CAS No. 96491-05-3

Expiry Date 03.09.2023
Lot Number 855474CY
Store at 20°C in the dark

Please note: The expiry date is valid under recommended storage conditions only.

Gravimetric Data			
Product Name	Conc. (mg/l)	Purity %	Weight (mg)
Thenylchlor	10.000	96.9	1.035
Solvent Information			
Solvent	Lot No.	Exact Quantity (ml)	
Cyclohexane	I093411717	100.00	
Traceability Data			
17445500 G979779	neat product		
17445500 855474CY	10.000 mg/l		
Analytical Data			
Detection: HPLC/DAD	Method Details:		
Column: ReproSil 100 C18 5µ 250x3	Acetonitrile:H ₂ O 4:1		
Inj.-Vol.: 10.00 µl			
Flow: 1.0 ml/min			
Ret.-Time: 2.19 min.			
Identity check UV, RT			
Comment Property value was determined by gravimetric measurements and confirmed by peak area comparison.			
The uncertainty/tolerance of this standard is +/- 2.0 %, calculated in accordance with the EURACHEM/CITAC Guide - Quantifying Uncertainty in Analytical Measurement - Second Edition. The uncertainty given is the expanded combined uncertainty and represents an estimated standard deviation equal to the positive square root of the total variance of the uncertainty of components. The expanded uncertainty is U which is U _c (y)*K, where K is the coverage factor at the 95% confidence level (K=2). The expanded uncertainty is based on the combination of uncertainties associated with each individual operation involved in the preparation of this product.			

Certified on 03.09.2018

by D. Schmid

The Laboratory LGC Labor GmbH is accredited by DAkkS as indicated by the Accreditation Number D-RM-19883-01 & D-PL-19883-01 has shown competence based on ISO Guide 34:2009 with relevant parts of DIN EN ISO/IEC 17025:2005 for production of certified reference materials in form of organic pure substances and in form of single and multi-component solutions organic pure substances.


LGC Labor GmbH · Bgm.-Schlosser-Str. 6 A · 86199 Augsburg · Germany
Phone +49 821 906080 · Fax +49 821 9060888 · augsburg.inquiry@lgcgroup.com
The warranty for this product is limited to the purchasing price of this product.

257

Data file: 17445500-11-r001.dx
Sample name: 855474CY
Inj. volume [µl]: 10.0
Acq. method: S1_41K.amx

Instrument: DAD5
Sequence Name: 31082018-KPK
Injection date: 9/1/2018 3:17:16 AM
Location: P3-A9

Sample Description: Thenylchlor



Signal: DAD1 Total WL (262.0:262.0 nm)

Nr.	RT [min]	Area	Height	Area%
1	2.19	97.87367	19.28	100.00
	Sum	97.87		

Schmid D