

Isotopic Standard

Lithium Isotope Standard

Product #: VHG-LIS6LIZ-100

Matrix: 2% HNO₃

Lot #: 1000801-11

Element	Concentration and Uncertainty		
⁶ Li	Certified W/V	100 µg/mL	± 1 µg/mL

Isotopic Composition						
Isotope	Abundance	Relative Std Error (%)		Isotope	Abundance	Relative Std Error (%)
⁶ Li	95.0	--		⁷ Li	5.0	--

This solution is intended for use as a calibration or reference standard for inductively coupled plasma mass spectrometry (ICP-MS) where known values for isotopic abundance are required.

Certification: VHG standards are manufactured and certified under a quality control system that is accredited to both **ISO 9001** and **ISO/IEC 17025**. This standard was prepared to a nominal concentration of 100µg/mL by gravimetric methods. Matrix components (see "Matrix") were of high purity grade reagents and all dilutions were done with filtered (0.22µm), 18 M-ohm deionized water. Certified concentrations are based upon gravimetric procedures. The uncertainty associated with the certified concentrations is +/- 1% relative, which is the sum of the estimated errors due to the purity of the raw materials, the gravimetric preparation of the solution and transpiration through the container wall. Atomic abundance and/or ratio values are based upon raw material values and are verified using ICP-MS analysis.

Tools: The balances used in the preparation of VHG CRMs are calibrated regularly with traceability to NIST. All volumetric dilutions are performed in class A glassware which is recalibrated regularly according to NIST recommended procedures. Measurement of standard preparation temperature was done using a calibrated thermometer maintained under internal procedure 4.11-A.

Recommendations: VHG guarantees the accuracy of this solution for **18 Months** from the certification date shown below, provided it is kept tightly capped and stored under normal laboratory conditions. We recommend that the analyst: (1) mix the solution by gentle shaking immediately prior to use, (2) use only pre-cleaned containers and transfer-ware, (3) make dilutions using certified volumetric class A flasks and pipettes, and, (4) dilute with the same matrix as the original standard.

VHG Labs, Inc.



Chuck Goudreau, Certifying Officer

See Exp. Date on Container
Certification Date

LGC waives all responsibility for any damages resulting from the usage and/or implementation of the products/data described herein.



VHG Custom Standards are Traceable to the Following NIST SRMs:

Analyte	Aq. SRM	MO SRM		Analyte	Aq. SRM	MO SRM		Analyte	Aq. SRM	MO SRM
Ag	3151	1077a		Hf	3122	—		S	3154	2770
Al	3101a	1075a		Hg	3133	3133		Sb	3102a	3102a
As	3103a	3103a		Ho	3123a	—		Sc	3148a	3148a
Au	3121	—		In	3124a	3124a		Se	3149	3149
B	3107	3107		K	3141a	3141a		Si	3150	1066a
Ba	3104a	1051b		La	3127a	3127a		Sm	3147a	—
Be	3105a	3105a		Li	3129a	3129a		Sn	3161a	1057b
Bi	3106	3106		Lu	3130a	—		SO ₄ ⁻²	3181	—
Br ⁻	3184	—		Mg	3131a	3131a		Sr	3153a	3153a
Ca	3109a	3109a		Mn	3132	3132		Ta	3155	—
Cd	3108	1053a		Mo	3134	3134		Tb	3157a	—
Ce	3110	3110		Na	3152a	1069b		Te	3156	—
Cl ⁻	3182	1818a		Nb	3137	—		Th	3159	—
Co	3113	3113		Nd	3135a	—		Ti	3162a	3162a
Cr	3112a	1078b		Ni	3136	1065b		Tl	3158	3158
Cs	3111a	—		NO ₃ ⁻	3185	—		Tm	3160a	—
Cu	3114	1080a		P	3139a	3139a		U	3164	—
Dy	3115a	—		Pb	3128	1059c		V	3165	1052b
Er	3116a	—		Pd	3138	—		W	3163	3163
Eu	3117a	—		PO ₄ ⁻³	3186	—		Y	3167a	3167a
F ⁻	3183	—		Pr	3142a	—		Yb	3166a	—
Fe	3126a	1079b		Pt	3140	3140		Zn	3168a	3168a
Ga	3119a	—		Rb	3145a	—		Zr	3169	3169
Gd	3118a	—		Re	3143	—				
Ge	3120a	—		Rh	3144	3144				