

Certificate of Analysis

General Laboratory Reagent

Sodium Hydroxide 10M (10N)

| | |
|-----------------------------|-------------------|
| <i>Product No:</i> | S10001 |
| <i>Lot No:</i> | S30WW14E1 |
| <i>Expiry date:</i> | 11/02/2016 |
| <i>Assay Value:</i> | 10.0 M |
| <i>Date of measurement:</i> | 22/05/2014 |

Specification:

9.6M - 10.4M

Method:

Tested by potentiometric titration, to in-house method TPATB

Traceability:

This volumetric solution was checked by means of Hydrochloric acid Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 723d Tris(hydroxymethyl)aminomethane. All raw materials used to prepare this product are of high purity.

Products are manufactured under an NSAI registered I.S EN ISO9001:2008 Quality System, registration no: 19.2769

Date of issue of the certificate: 22/05/2014



Q.C. Supervisor Margaret Dornan

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-14E4