



Certificate of Analysis

Primary Turbidimetric Calibration Standard

0.5 NTU Ratio Turbidity Standard

Product No:	CRSR-0.5-500
Lot No:	CRSR0513L1
Expiry date:	28/11/2015

Certified value:	0.50 NTU
Date of measurement:	02/12/2013

Assay Limits:
0.45- 0.55 NTU

Method:
Tested by turbidity measurement to in-house method TPTURB

Traceability:
This turbidity standard is directly traceable to Standard Reference Material of National Institute of Standards and Technology (USA), 1690 and 1691.

Reference:
Water quality-Determination of turbidity. International Standard. ISO 7027 (1990), DIN 38 404 and NF EN 27027. Second Edition.

Products are manufactured under an NSAI registered I.S EN ISO9001:2008 Quality System, registration no: 19.2769

Date of issue of the certificate: 02/12/2013

A handwritten signature in black ink, appearing to read 'Margaret Dornan'.

Q.C. Supervisor Margaret Dornan

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-13L2

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com