

Certificate of Analysis



ISO 17034 Reference Material

Product Identification

Article Code: DRE-C10100500
Article Name: D-Allethrin
Formula: C₁₉H₂₆O₃
Mol. Weight: 302.41
CAS No.: 231937-89-6

Lot Number: G988236
Expiry Date: 18.12.2024
Storage Temperature: 4°C ± 4°C

Storage and handling: The RM should be stored in the original sealed bottle at the temperature given above. After use the bottle should be tightly closed and protected from moisture.

Purity:	94.19% (g/g)
Expanded Uncertainty U=	0.57% (g/g)

The uncertainty of this standard is calculated in accordance with the ISO 17034 and EURACHEM/CITAC Guide - Quantifying Uncertainty in Analytical Measurement, Second Edition. The expanded uncertainty is $U(\text{exp}) = u(\text{RM}) \times k$, where k is the coverage factor at the 95% confidence level ($k=2$). Uncertainty $u(\text{RM})$ is based on the combination of the uncertainties associated with each individual operation involved in the analysis of the product: $u(\text{RM}) = \sqrt{u(\text{char})^2 + u(\text{bb})^2 + u(\text{Its})^2 + u(\text{sts})^2}$; $u(\text{char})$ is the uncertainty of characterisation; $u(\text{bb})$ uncertainty of homogeneity test; $u(\text{Its})$ uncertainty of stability test long-term; $u(\text{sts})$ uncertainty of stability test short-term. $u(\text{Its})$ and $u(\text{sts})$ are not included in the calculation as the stability statement is based on real evidence opposed to simulation.

Minimum sample: 1 mg is recommended as the minimal sample amount. If less material is used, it is recommended to increase the certified uncertainty by a factor of two for half sample and a factor of four for a quarter of sample.

Intended use: Use this RM as calibrant for chromatography or any other analytical technique.

Analytical Data

Traceability of chromatography: To the International System of Units (SI).

Instrument:	GC/FID	Injector:	320°C
Detection:	FID	Initial Temp:	120°C for 4 min
Column:	Optima-SMS, 0.25 µm, 0.25 mm	End Temp:	320°C for 3 min
Inj.-Vol.:	2 µl	Gradient:	15°C/min
Flow:	1.0 ml/min		
Ret.Time:			

Comment

Traceability: The balances used are calibrated with weights traceable to the national standards (DKD).

Calibrated class A glassware is used for volumetric measurements.

Water Content: 0.24% (g/g) by Karl-Fischer-Titration ($U(\text{exp}) = 0.07\%$ (g/g)).

Purity was determined by chromatographic assay, corrected by water content and/or residue solvents.

Identity: EA, NMR, RT, IR, MS

No chromatogram available.

Mixture of Isomers. Ratio unknown.

Certificate Revision 1 - 18.12.2018 - N. Müller

Certified on: 18.12.2018
Certified by: N. Müller
RM Release

The LGC Labor GmbH, accredited by DAkkS as indicated by the accreditation number D-RM-19883-01 & D-PL-19883-01, has shown competence based on ISO 17034:2017 with relevant parts of DIN EN ISO/IEC 17025:2018 for production of certified reference materials in form of organic pure substances and in form of single and multi-component solutions of organic pure substances.

LGC Labor GmbH - Bgm.-Schlosser-Straße 6A - 86199 Augsburg - Germany
Phone +49 821 906080 - Fax +49 821 9060888 - augsburg.inquiry@lgcgroup.com
The warranty for this product is limited to the purchasing price of this product.