



Shannon Free Zone, Shannon, Co. Clare, Ireland
Tel: +353 61 472622 Fax: +353 61 472642
Email: sales@reagecon.ie
www.reagecon.com

CERTIFICATE OF ANALYSIS

PRODUCT: Tellurium 1000 \pm 10mg/l in 2 - 5% HNO₃

PRODUCT CODE: PTE2A10

LOT NUMBER: PTE21015E1

DATE OF MEASUREMENT: 05th May 2015

EXPIRY DATE: 28th May 2017

MEAN RESULT: 1000mg/l Te by ICP

SPECIFICATION: 990 - 1010mg/l

TRACE METALLIC IMPURITIES/ STARTING MATERIAL PURITIES:

Higher than 100ppb: NONE detected by ICP-MS. (Except Si - < 1mg/L)

PREPARATION: This standard was prepared gravimetrically to contain the elemental concentration as shown. Raw materials of the highest purity were used, including high purity acids (where applicable), and 18 megohm water. Equipment used; calibrated Class A glassware and balances traceable to National Standards. All bottles were acid leached and triple rinsed with double deionized water prior to use.

TRACEABILITY: This standard is certified traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 3156 Tellurium Standard Solution.

Date: 05th May 2015

This certificate must not be reproduced except in full.