

REFERENCE MATERIAL CERTIFICATE

MIX 9 (AFLATOXINS)

1. General information

This document is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide 31 [1] and Eurachem / CITAC Guides [2,3].

2. Description of the Reference Material (RM)

Name:	Mix 9 (Aflatoxins)
CAS number:	10002879
Catalog number:	DRE-V30000006AL
Lot #:	1000002265
Certificate version:	1
Expiry date:	15.07.2021
Starting material 1:	Aflatoxin B1 Lot # S16251B, Aflatoxin B2, Lot #0404049 Aflatoxin G1, Lot # S18255G, Aflatoxin G2 Lot #0504071 Romer Labs Diagnostic GmbH
Physical description of RM:	Solution of 4 different Aflatoxins in acetonitrile
Packaging and amount of RM:	<u>DRE-V30000006AL</u> : Amber glass ampoules fitted with teflon faced butyl septa and PP screw caps, solution of 5 mL
Name and address of the manufacturer:	Romer Labs Diagnostic GmbH Technopark 5, 3430 Tulln, Austria www.romerlabs.com
Name and address of the supplier:	LGC Standards GmbH Mercatorstraße 51, 46485 Wesel, Germany Tel +49(0)2 81 98 87 0, Fax +49(0)2 81/98 87 199 www.lgcstandards.com

2.1 Intended use of the RM

- for laboratory use only
- calibration of analytical instruments

2.2 Instruction for the correct use of the RM

The ampoules should be stored at -18 to -22°C in a dark place. Before usage of the RM, the ampoules should be allowed to warm to room temperature. The recommended minimum sub-sample amount for all kinds of application is 100 µL. The expiry date of this RM is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/packages.

2.3 Hazardous situation

The normal laboratory safety precautions should be observed when working with this RM. Further details for the handling of this RM are available as safety data sheet (SDS).

Hazardous Ingredients	Concentration in %	Pictograms	Signal word	Hazard statement(s)
Acetonitrile	> 99.9		Danger	H225, H302, H312, H319, H332

REFERENCE MATERIAL CERTIFICATE

3. Certified values and their uncertainties

Mix 9 (Aflatoxins)		
Compound	Mass concentration ^a	
	Certified value ^b	Uncertainty ^c
Aflatoxin B1	1.02 µg/mL	± 0.01 µg/mL
Aflatoxin B2	1.01 µg/mL	± 0.01 µg/mL
Aflatoxin G1	1.02 µg/mL	± 0.01 µg/mL
Aflatoxin G2	1.04 µg/mL	± 0.01 µg/mL

^a Values are based on preparation data and confirmed experimentally by HPLC-UV
^b Mass concentration based on weighed amount, purity and dilution steps
^c Expanded uncertainty U (k = 2) of the value u_c according to GUM [4]

3.1 Calculation of uncertainty

The uncertainty of the calibrant solution was calculated on the basis of preparation [5].

Uncertainty components	Description	Standard uncertainty (u)	
Purity (P) of solid Aflatoxin B1 Aflatoxin G1 Aflatoxin B2 Aflatoxin G2	P ₁ = 99 ± 1 % P ₂ = 99 ± 1 % P ₃ = 99 ± 1 % P ₄ = 99 ± 1 %	u (P₁₋₄) = 0.6 %	a
Weighing procedure weighted sample: m _{wsAFB1} = 2.577 mg m _{wsAFB2} = 2.546 mg m _{wsAFG1} = 2.578 mg m _{wsAFG2} = 2.627 mg	U(m) = 0.0026 mg + 9.51 * 10 ⁻⁶ * m _{Toxin} u(m) = U(m)/2	u (m) = 0.0013 mg	b
Aflatoxin B1 + B2 + G1 + G2 stock solution: volumetric flask 1: V _{f1} = 250 mL Aflatoxin mixture working solution: volumetric flask 2: V _{f2} = 500 mL one-mark glass pipette: V _p = 50 mL	calibration flask 1: 250 mL ± 0.15 mL repeatability flask 1: 0.03 mL volume expansion solvent flask 1 calibration flask 2: 500 mL ± 0.25 mL repeatability flask 2: 0.1 mL volume expansion solvent flask 2 calibration pipette: 50 mL ± 0.075 mL volume expansion solvent pipette	u (cal1) = 0.06 mL u (rep1) = 0.03 mL u (Vol. exp.1) = 0.59 mL u (V1) = 0.6 mL u (cal2) = 0.1 mL u (rep2) = 0.1 mL u (Vol. exp.2) = 1.19 mL u (V2) = 1.2 mL u (cal3) = 0.03 mL u (Vol. exp.3) = 0.12 mL u (V3) = 0.1 mL	c d e f g h i j k l m

^a Maximum tolerance of purity was divided by $\sqrt{3}$
^b Calculation of this u-value is based upon the uncertainty formula for the weighed amount as given in the calibration report from annual balance calibration
^{c,g,k} A triangular distribution (division by $\sqrt{6}$) was chosen for the calculation of u (cal)
^{d,h} Based on a series of ten fill and weigh experiments on a typical 250 mL and 500 mL flask; the value was used directly as a standard deviation
^{e,l} Based on the density of 0.7857 g/cm³ at temperature T = 20°C and a maximum temperature variation of ± 3°C, of volume expansion, relative volume expansion coefficient of acetonitrile is 1370 * 10⁻⁶/°C [6], volume expansion term (rectangular distribution) was divided by $\sqrt{3}$
^{f,j,m} All contributions are combined to give the u (V) = $\sqrt{u(\text{cal})^2 + u(\text{rep})^2 + u(\text{Vol. exp.})^2}$

REFERENCE MATERIAL CERTIFICATE

Calculation of the combined uncertainty u_c and the expanded standard uncertainty U for Aflatoxin B1 as example

$$c_{\text{Toxin}} = \frac{10 \times m_{\text{WSAFB1}} \times P_1 \times V_p}{V_{f1} \times V_{f2}} = \frac{10 \times 2.577 \times 99 \times 50}{250 \times 500} = 1.02 \text{ mg/L}$$

$$\frac{u_c(c_{\text{Toxin}})}{c_{\text{Toxin}}} = \sqrt{\left[\frac{u(P_1)}{P_1}\right]^2 + \left[\frac{u(m)}{m_{\text{WSAFB1}}}\right]^2 + \left[\frac{u(V1)}{V_{f1}}\right]^2 + \left[\frac{u(V2)}{V_{f2}}\right]^2 + \left[\frac{u(V3)}{V_p}\right]^2} = \sqrt{\left[\frac{0.61}{99}\right]^2 + \left[\frac{0.00131}{2.577}\right]^2 + \left[\frac{0.6}{250}\right]^2 + \left[\frac{1.2}{500}\right]^2 + \left[\frac{0.1}{50}\right]^2} = 0.007$$

$$u_c(c_{\text{Toxin}}) = c_{\text{Toxin}} \times 0.007 = 1.02 \times 0.007 = 0.007 \text{ mg/L}$$

Calculation of expanded standard uncertainty U using a coverage factor $k = 2$


$$U(c_{\text{Toxin}}) = u_c(c_{\text{Toxin}}) \times 2 = 0.007 \times 2 = 0.014 \text{ mg/L} = 0.01 \text{ } \mu\text{g/mL}$$

4. Discussion of traceability

This calibrant is certified on the basis of gravimetric preparation [5]. Thus the certified values (mass concentrations of 4 different Aflatoxins) are based on the weighed amount of the starting materials and are therefore traceable to the stated purity of the solid raw materials. High purity materials represent a practical realization of concentration units, through conversion of mass to molar quantity.

6. Confirmation of certified value by HPLC-UV

The certified concentration of Mix 9 of the gravimetric prepared solution was confirmed by HPLC-UV against an independently prepared reference batch of unlabeled Aflatoxin (B1, B2, G1, G2) calibrants.

<table style="width: 100%; border-collapse: collapse;"> <tr><td>column</td><td colspan="3">Phenomenex Luna C18(2), 250 x 3.0 mm, 5μ</td></tr> <tr><td>injection volume</td><td colspan="3">100 μL sample</td></tr> <tr><td>solvent A</td><td colspan="3">water / acetonitrile / methanol 57/17/26</td></tr> <tr><td>oven</td><td colspan="3">30°C</td></tr> <tr><td>flow rate</td><td colspan="3">0.5 mL / min</td></tr> <tr><td>DAD settings</td><td colspan="3">365 nm</td></tr> <tr><td>sample dilution</td><td colspan="3">1:5 with water</td></tr> <tr><td></td><td style="text-align: center;">Time [min]</td><td style="text-align: center;">area</td><td style="text-align: center;">concentration ^a</td></tr> <tr><td>Aflatoxin G2</td><td style="text-align: center;">8.220</td><td style="text-align: center;">2.449</td><td style="text-align: center;">1.04 \pm 0.02 μg/mL</td></tr> <tr><td>Aflatoxin G1</td><td style="text-align: center;">9.853</td><td style="text-align: center;">2.184</td><td style="text-align: center;">1.01 \pm 0.02 μg/mL</td></tr> <tr><td>Aflatoxin B2</td><td style="text-align: center;">11.147</td><td style="text-align: center;">3.132</td><td style="text-align: center;">1.00 \pm 0.02 μg/mL</td></tr> <tr><td>Aflatoxin B1</td><td style="text-align: center;">13.567</td><td style="text-align: center;">2.829</td><td style="text-align: center;">1.02 \pm 0.02 μg/mL</td></tr> </table> <p>^a Mean of 6 replicate measurements against reference batch, confidence interval with P = 95 %</p>	column	Phenomenex Luna C18(2), 250 x 3.0 mm, 5 μ			injection volume	100 μ L sample			solvent A	water / acetonitrile / methanol 57/17/26			oven	30°C			flow rate	0.5 mL / min			DAD settings	365 nm			sample dilution	1:5 with water				Time [min]	area	concentration ^a	Aflatoxin G2	8.220	2.449	1.04 \pm 0.02 μ g/mL	Aflatoxin G1	9.853	2.184	1.01 \pm 0.02 μ g/mL	Aflatoxin B2	11.147	3.132	1.00 \pm 0.02 μ g/mL	Aflatoxin B1	13.567	2.829	1.02 \pm 0.02 μ g/mL	 <p style="text-align: center;"><i>Figure 1: HPLC-UV chromatogram of Mix 9</i></p>
column	Phenomenex Luna C18(2), 250 x 3.0 mm, 5 μ																																																
injection volume	100 μ L sample																																																
solvent A	water / acetonitrile / methanol 57/17/26																																																
oven	30°C																																																
flow rate	0.5 mL / min																																																
DAD settings	365 nm																																																
sample dilution	1:5 with water																																																
	Time [min]	area	concentration ^a																																														
Aflatoxin G2	8.220	2.449	1.04 \pm 0.02 μ g/mL																																														
Aflatoxin G1	9.853	2.184	1.01 \pm 0.02 μ g/mL																																														
Aflatoxin B2	11.147	3.132	1.00 \pm 0.02 μ g/mL																																														
Aflatoxin B1	13.567	2.829	1.02 \pm 0.02 μ g/mL																																														

REFERENCE MATERIAL CERTIFICATE

7. Further information

The purchaser must determine the suitability of this product for its particular use. LGC Standards GmbH makes no warranty of any kind, express or implied, other than its products meet all quality control standards set by LGC Standards GmbH. We do not guarantee that the product can be used for a special application.

approved for release by: *Laurence Treccani-Chinelli, Global Supply Chain Manager - LGC Standards* date: 21.01.2020

This document has been computer generated and is valid without a signature.

References:

- [1] ISO Guide 31:2015 - 1-18, "Reference materials – contents of certificates, labels and accompanying documentation"
- [2] Eurachem / CITAC Guide, 1-37, (2003), "Traceability in Chemical Measurement"
- [3] Eurachem / CITAC Guide CG4, 1-133, (QUAM:2012.P1), "Quantifying Uncertainty in Analytical Measurement", 3rd Ed.
- [4] International Organization for Standardization (ISO), (2008), "Guide to the expression of uncertainty in measurement", (GUM 1995 with minor corrections) 1st Ed. Geneva, Switzerland
- [5] R.D. Josephs, R. Krska, S. MacDonald, P. Wilson, H. Pettersson, J. AOAC Int. **86**, 50-60, (2003), "Preparation of a Calibrant as Certified Reference Material for Determination of the Fusarium Mycotoxin Zearalenone"
- [6] E.W. Flick, (1998), "Industrial Solvents Handbook", 5th Ed., Noyes Data Corp. Westwood NJ