


# Certificate of Analysis

## Androsterone-D<sub>4</sub> glucuronide

Cerilliant Quality
ISO GUIDE 34
ISO/IEC 17025
ISO 13485
ISO 9001
GMP/GLP

**Catalog Number:** A-099  
**Solution Lot:** FE011312-01  
**Retest Date:** February 2015  
**Solvent:** Methanol  
**Volume per Ampule:** Not less than 1 mL  
**Storage:** Store unopened in freezer.  
**Shipping:** Ship cold. See Stability Section.  
**Intended Use:** For R&D/ analytical purposes only. Not suitable for human or animal consumption.  
**Regulatory:** Canadian TK# 61-671



**Safety:** Flammable, Poison

- Retest Date - stability studies ongoing. Certificate of Analysis will be updated upon completion of retest.
- Ampoules are overfilled to ensure a minimum 1 mL volume fill. We advise laboratories to use measured volumes of this standard solution before diluting to the desired concentration.

Component	Solution Purity	Certified Concentration
Androsterone-D <sub>4</sub> glucuronide	99.6%	100.0 ± 0.9 µg/mL
<ul style="list-style-type: none"> <li>Uncertainty of the concentration is expressed as an expanded uncertainty in accordance with ISO 17025 and Guide 34 at the approximate 95% confidence interval using a coverage factor of <math>k = 2</math> and has been calculated by statistical analysis of our production system and incorporates uncertainty of the purity factor, material density, and balance and weighing technique.</li> <li>This standard is prepared gravimetrically and mass results are reported on the conventional basis for weighing in air. Concentration is calculated based on: the actual measured mass; Purity Factor of the analyte(s); measured mass of the solution; and the density of the pure diluent at 20°C.</li> <li>Concentration is corrected for chromatographic purity, residual water, residual solvents and residual inorganics.</li> </ul>		

### Solution Standard Verification and Homogeneity

Standard Solution	Lot Number	Verified Concentration (µg/mL)		%RSD - Homogeneity	
		Actual Results	Acceptance Criteria	Actual Results	Acceptance Criteria
New Lot	FE011312-01	99.7	± 5%	1.1	≤ 3%
<ul style="list-style-type: none"> <li>Concentration is verified through multiple analyses and is calculated as the average of multiple analyses compared to an independently prepared calibration solution.</li> <li>Homogeneity of the New Lot is ensured through rigorous production process controls statistically analyzed to evaluate risk and verified by analysis. The % RSD of samples pulled from across the lot demonstrate homogeneity of the New Lot.</li> </ul>					

### Traceability

- Gravimetrically prepared using qualified balances calibrated semi-annually by Mettler Toledo using NIST traceable weights. Calibration verification performed weekly and prior to each use utilizing NIST traceable weights. Each balance has been assigned a minimum weighing by Mettler Toledo taking into consideration the balance and installed environmental conditions to ensure weighing complies with USP tolerances of no more than 0.1% relative error.
- Concentration is verified against an independently prepared calibration solution gravimetrically prepared using balances calibrated to NIST.
- In addition, each neat material utilized has been identified and thoroughly characterized through the use of multiple analytical techniques. Spectral data is provided on subsequent pages of the COA.

Cerilliant certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration/retest date when stored unopened as recommended. Product should be used shortly after opening to avoid concentration changes due to evaporation. Warranty does not apply to ampoules stored after opening.




Darron Ellsworth, Quality Assurance Manager

April 15, 2014

Date

## Standard Solution Assay Parameters

**Analysis Method:** LC/MS  
**Column:** Luna 3 $\mu$  C18, 2.0 x 100 mm  
**Mobile Phase:** Acetonitrile::0.1% Formic acid (40::60)  
**Flow Rate:** 0.3 mL/min  
**Ionization:** Electrospray, Negative Ion

## Calibration Curve

**Calibration Curve:** Linear Regression  
**Number of Points:** 3  
**Linearity (r):** 0.990

## Neat Material Data

**Compound Name:** Androsterone-D<sub>4</sub> glucuronide  
**Compound Lot:** FN120911-02

**Chemical Formula:** C<sub>23</sub>H<sub>34</sub>D<sub>4</sub>O<sub>8</sub>  
**CAS Number:** NA  
**Molecular Weight:** 470.59

### Neat Material Characterization Summary


Analytical Test	Method	Results
Chromatographic Purity by LC/MS Analysis	SP10-0107	> 99.9%
Identity by LC/MS Analysis	SP10-0107	Consistent with Structure
Isotopic Purity by LC/MS SIM Analysis	SP10-0107	0.03% D <sub>0</sub> vs D <sub>4</sub>
		0.02% D <sub>0</sub> 21.92% D <sub>3</sub>
		0.16% D <sub>1</sub> 75.28% D <sub>4</sub>
		2.62% D <sub>2</sub>
Identity by <sup>1</sup> H-NMR Analysis	USP <761>, SP10-0116	Consistent with Structure
Residual Solvent Analysis by GC/FID Headspace	AM1087 <sup>1</sup>	1.42%
Residual Water Analysis by Karl Fischer Coulometry	USP <921>, SP10-0103	0.44%
Inorganic Content by Microash Analysis	SP10-0135	< 0.2%
Assay by Quantitative NMR Analysis	USP <761>, SP10-0116	90.33%
Purity Factor		90.33%

- ♦ Chromatographic purity is calculated as the average of two independently performed analyses utilizing two different methods. Acceptance criteria requires the purity values to be within 0.5% of each other.

<sup>1</sup> Validated analytical method


## Spectral and Physical Data

### LC/MS



**Column:** Zorbax Eclipse Plus C18 RRHD 1.8 $\mu$ , 2.1 x 50 mm  
**Mobile Phase:** 0.1% Formic acid in Water::Acetonitrile  
(80::20 to 40::60 at 2.5 mins, to 80::20 at 4.1 mins,  
hold 3.9 mins)  
**Flow Rate:** 0.4 mL/min  
**Scan Range:** 50-1200 amu  
**Ionization:** Electrospray, Negative Ion

**Data File Name:** W12281117  
**Instrument:** Waters XEVO G2 QTOF  
**Sample Name:** FN120911-02  
**Acquired:** December 28, 2011




## Spectral and Physical Data (cont.)

### Isotopic Purity by LC/MS SIM Analysis


**Column:** Zorbax Eclipse Plus C18 RRHD 1.8 $\mu$ , 2.1 x 50 mm  
**Mobile Phase:** 0.1% Formic acid in Water::Acetonitrile  
(80::20 to 40::60 at 2.5 mins, to 80::20 at 4.1 mins,  
hold 3.9 mins)  
**Flow Rate:** 0.4 mL/min  
**Scan Range:** 465-469 amu  
**Ionization:** Electrospray, Negative Ion

**Data File Name:** W12281117  
**Operator:** JDC  
**Instrument:** Waters XEVO G2 QTOF  
**Sample Name:** FN120911-02  
**Method File:** 28-64C3  
**Acquired:** December 28, 2011



**Spectral and Physical Data (cont.)**

**Residual Solvent Analysis by GC/FID Headspace**




**Column:** DB-ALC1 30 m x 0.53 mm, 3 µm film thickness  
**Temp Program:** 40°C (12 min) to 220°C at 40°C/min (5.5 min)  
**Carrier Gas:** Helium  
**Flow Rate:** 2.0 mL/min  
**Detector Heater Temp:** 250°C  
**Injector:** Headspace Sampler  
**HS Oven Temp:** 60°C  
**Vial Equilibration:** 10 minutes

**Data File Name:** C:\CHEM32\1\DATA\GC91211\1222114.D  
**Instrument:** GC#9  
**Sample Name:** FN120911-02  
**Acquired:** December 22, 2011

Peak	Compound	Area	Weight %
1	Methanol	194.41	1.35
2	Acetonitrile	15.22	0.07
3	NMP	NA	NA
<b>Total</b>			<b>1.42</b>

**<sup>1</sup>H NMR**

**Instrument:** JEOL ECS 400  
**Solvent:** DMSO-D<sub>6</sub>



## Stability

<b>Short Term Stability :</b> A summary of accelerated stability findings for this product is listed below.		
Storage Condition	Mean Kinetic Temperature (MKT)	Time Period
Freezer	-15°C	No decrease in purity was noted after one week.
Refrigerator	4°C	No decrease in purity was noted after one week.
Room Temperature	21°C	No decrease in purity was noted after one week.
40°C	40°C	~ 1% decrease in purity was noted after one week
<b>Transport/Shipping :</b> Stability data supports transport of this product at MKT not exceeding 4°C over a period of one week.		
<b>Short Term Storage:</b> Stability data supports short term storage up to 3 months at Refrigerate conditions.		
<b>Long Term Stability:</b> Long term stability has been assessed for freezer storage (-10 °C to -25 °C) conditions. Stability of a minimum of 25 months has been established through real-time stability studies.		

## COA Revision History

Revision No.	Date	Reason for Revision
00	February 3, 2012	Initial version
01	February 6, 2013	Revised Retest Date from March 2013 to March 2014.
02	April 15, 2014	Revised Retest Date from March 2014 to February 2015.