

Rio Tinto Alcan Inc. - CRDA / ARDC

1955, boul. Mellon Éd. - BLDG. 110

C.P. 1250

Québec, Canada, G7S 4K8

Tel.: (418) 699-6585 Fax: (418) 699-2919



Certificat d'analyse
Matériau de référence certifié
Certificate of analysis
Spectrochemical Standard

6066 AF

Certifié le / Certification: 05/12/2006

Élément	Element	Si	Fe	Cu	Mn	Mg	Cr	Ni	Zn
% (p/p)	Weight % (w/w)	0.97	0.60	0.96	0.71	0.97	0.018	0.032	0.28
Incertitude	Uncertainty	± 0.009	± 0.010	± 0.011	± 0.006	± 0.012	± 0.0009	± 0.0009	± 0.006

Élément	Element	Ti	As	B	Be	Bi	Ca	Cd	Co
% (p/p)	Weight % (w/w)	0.032			0.0013	0.007	0.0028	0.0011	
Incertitude	Uncertainty	± 0.0016			± 0.00008	± 0.0010	± 0.00029	± 0.00008	

Élément	Element	Ga	Hg	K	Li	Mo	Na	P	Pb
% (p/p)	Weight % (w/w)	0.021			0.0034i		0.0039i	0.0032	0.0075
Incertitude	Uncertainty	± 0.0007			± 0.00013		± 0.00010	± 0.0014	± 0.00027

Élément	Element	Sb	Sn	Sr	V	Zr
% (p/p)	Weight % (w/w)	0.0026	0.014	0.0029	0.021	0.0032
Incertitude	Uncertainty	± 0.00044	± 0.0008	± 0.00051	± 0.0006	± 0.00013

("i" valeur individuelle, "L" plus petite que, "N" valeur non certifiée)

("i" individual value, "L" less than, "N" value not certified)

Par la présente nous certifions les analyses ci-dessus
We hereby certify the above analysis

Hafida Hamouche