

Certificate of Analysis

Product name:	Urine-Calibrator, lyophil., for Catecholamines
Order no.:	2013
Lot no.:	1126
Expiration date:	03/2019
Storage Temperature:	2-8°C

Date of QC - release: 28.04.16

Results of QC:

Lyophilisation:

- Lyophilisation process according to fixed program; no irregularities.
- Lyophilised material (pellets) optically o.k.

Reconstitution:

- Completely soluble in 8 ml of 0.2 M HCl.

Purity / Peakshape:

- symmetry within QC specs.

Mean values established:

- difference between spiked amounts and established mean values within QC specs.

Functional assay with the other kit components:

- Recovery within QC specs.
- Control values found within the given ranges
- No interfering peaks from matrix

Infectiousity:

- The material is made of human urine. Therefore the material has to be considered as potentially infectious and should be treated with appropriate care.

Munich, 10.08.2016

Iris Wendler
(QC-Manager)

This document has been produced electronically and bears no signature.